

Test Certificate

concerning the test of influences on MRI systems caused by
non-Siemens products

1. Test certificate, product designation and manufacturer

This is to certify that the following product

Penlon 451 MRI Anaesthesia System

of the manufacturer:

Penlon Limited
Abingdon Science Park
Barton Lane, Abingdon
OX14 3NB, UK

does not impair the functioning of the MRI systems of Siemens Healthcare GmbH, Magnetic Resonance Systems, specified in section 2:

Potential functional restrictions of the MRI systems are indicated in section 4.

Possible adverse effects of the MRI systems on the product indicated above are expressly not subject of the test on which this certificate is based. Therefore, this certificate does not imply non-interference of the above-mentioned product through the MRI system. The functional restrictions of the 3rd party product are indicated in section 4 and 5 have to be added to the operator manual of the 3rd party product.

2. 7T MAGNETOM systems involved

- 1 MAGNETOM Terra
- 2 MAGNETOM Terra.X

3. Applications

The “**Prima 451 MRI**” is intended for use on MR Systems till/including a Magnet field intensity of 7 Tesla.

4. Limitations

During the use of this product the following limitations of functionality and/or applications of the MRI systems, options and accessories apply.

- none

5. Warnings

- Magnetic attraction was observed at the 100mT line. The device must be placed carefully, and the brakes must be activated.

6. Validity

This test certificate shall be valid until revoked by Siemens Healthineers AG.

Siemens Healthineers AG



*Electronically signed by: Joerg Teiche
Reason: I.V.
Date: Feb 28, 2025 15:52 GMT+1*

Joerg Teiche Vice President Quality MR



*Electronically signed by: Florian Meise
Reason: I.V.
Date: Mar 3, 2025 08:17 GMT+1*

Dr. Florian Meise Vice President MR Portfolio, Partners and Digital