

25 July 2017

Dear Customer,

**Penlon Anaesthesia System - Validity for use in an MRI environment**

The following equipment has been tested for attraction, stability of performance and effect on image in close proximity to a Magnetic Resonance Imaging Machine.

Products	1.5T		3T	
	400 gauss	1000 gauss	400 gauss	1000 gauss
Prima 451 MRI Anaesthetic System	✓	✓	✓	✓
A200SP Absorber	✓	✓	✓	✓
High Suction Controller and collection jar	✓	✓	✓	✓
AV-S MRI Ventilator	✓	✓	✓	✓
Gauss Meter	✓	✓	✓	✓
Nuffield 200 Ventilator	✓	✓	✓	✓
Airway Pressure Alarm	✓	✓	✓	✓
Sigma Delta Vaporizer	✓	✓	✓	✓
MRI Breathing Circuits	✓	✓	✓	✓
AGSS Receiver	✓	✓	✓	✓
Diamond MRI Laryngoscope Range	✓	✓	✓	✓
Oxygen Flowmeter	✓	✓	✓	✓

The above products are validated for use in proximity to the 400 and 1000 gauss lines generated by 1.5T and 3T magnets if used in compliance with user instructions. Users should, as a precautionary measure, make an individual check of MRI suitability of any device with their particular MRI facility before use.

Please contact your Penlon representative if you have any questions.

Sincerely,



Roger Crossley  
Product Manager